

KPL LumiGLO® Chemiluminescent Substrate

<u>Catalog No.</u>	<u>Size</u>
5430-0042 (54-61-02)	60 mL
5430-0040 (54-61-00)	240 mL
5430-0041 (54-61-01)	720 mL

DESCRIPTION

KPL LumiGLO® is a luminol-based chemiluminescent substrate designed for use with peroxidase-labeled (HRP) reporter molecules. KPL LumiGLO provides increased sensitivity over chromogenic substrates in both blotting and microwell assays. Positive reaction sites are rapidly detected with high sensitivity and minimal background. In blotting applications, permanent results are recorded on X-ray film. The use of KPL LumiGLO allows for multiple stripping and reprobing of blots. In microwell assays, positive reactions are rapidly detected and read in a Luminometer. KPL LumiGLO provides a dynamic range that is linear for a longer period of time than other chemiluminescent substrates.

In the presence of hydrogen peroxide, HRP converts luminol to an excited intermediate dianion. This dianion emits light on return to its ground state. After reaction with HRP the light emission from KPL LumiGLO reaches maximum intensity within 5 minutes and is sustained for approximately 1–2 hours.

CONTENT

- 5430-0042 (54-61-02) contains:
 - 1 x 30 mL KPL LumiGLO Substrate A
 - 1 x 30 mL KPL LumiGLO Substrate B
 Sufficient material is supplied to process approximately 600cm² of membrane.
- 5430-0040 (54-61-00) contains:
 - 1 x 120 mL KPL LumiGLO Substrate A
 - 1 x 120 mL KPL LumiGLO Substrate B
 Sufficient material is supplied to process approximately 2400cm² of membrane.
- 5430-0041 (54-61-01) contains:
 - 3 x 120 mL KPL LumiGLO Substrate A
 - 3 x 120 mL KPL LumiGLO Substrate B
 Sufficient material is supplied to process approximately 7200cm² of membrane.

STORAGE/STABILITY

KPL LumiGLO is supplied as a two component system. Store at 2-8 °C. Stable for a minimum of one year from date of receipt when stored at 2-8 °C.

APPLICATIONS

KPL LumiGLO can be used in both microwell and blotting applications such as ELISA, Western blotting, Southern blotting, dot blotting, plaque and colony hybridizations.

SUGGESTED REAGENTS NOT INCLUDED

1. Primary antibody or DNA probe.
2. HRP-labeled antibody or streptavidin.
3. X-ray film.
4. Nylon, nitrocellulose or PVDF membrane.
5. Blocking Solution (See RELATED PRODUCTS).
6. Wash Solution (See RELATED PRODUCTS).
7. 20X SSC (See RELATED PRODUCTS).
8. KPL Protein Detector LumiGLO Western Blotting Kit provides HRP secondary antibodies, blocking solution, wash solution and KPL LumiGLO Chemiluminescent Substrate (See RELATED PRODUCTS).
9. KPL HRP Chemiluminescent Blotting Kit (See RELATED PRODUCTS).

PREPARATION

Mix Substrate A and Substrate B in equal volumes. Warm to room temperature before use. Solution need not be protected from light. Solution is stable for up to one hour at room temperature or up to 24 hours when stored at 2-8 °C.

BLOTTING PROCEDURES

All steps are at room temperature unless otherwise noted.

Note: KPL LumiGLO can be used with nitrocellulose, nylon and PVDF membranes.

Note: Milk or casein-based blocking solutions are recommended for use with KPL LumiGLO. BSA or serum-based blocking agents may cause elevated background. SeraCare recommends KPL Detector Block (See RELATED PRODUCTS) for the highest sensitivity with low background.



**KPL LumiGLO®
Chemiluminescent Substrate**

<u>Catalog No.</u>	<u>Size</u>
5430-0042 (54-61-02)	60 mL
5430-0040 (54-61-00)	240 mL
5430-0041 (54-61-01)	720 mL

WESTERN BLOTTING

Detection:

1. Perform gel electrophoresis and transfer following standard procedures.
2. Block the membrane with KPL Detector Block (See RELATED PRODUCTS) or other appropriate blocking solution, for 1 hour at room temperature or overnight at 2- 8°C.
3. Incubate membrane with primary antibody or serum sample, diluted in blocking solution, for 1 hour.
4. Wash membrane with KPL Wash Solution (See RELATED PRODUCTS) or other appropriate wash solution, 3 times for 5 minutes each.
5. Incubate with HRP conjugate, diluted in blocking solution, for 1 hour. The concentration of HRP conjugate must be determined experimentally.
6. Wash 3 times for 5 minutes each.
7. Prepare KPL LumiGLO Chemiluminescent Substrate by mixing equal volumes of Substrates A and B. Incubate membrane for 1 minute in the KPL LumiGLO working solution (use approximately 1 mL per 10 cm² membrane).
8. Remove membrane from KPL LumiGLO and touch the corner to a piece of filter paper. Place membrane between plastic sheets or in a hybridization bag.
9. Expose membrane to X-ray film and indicator. The signal obtained from the first exposure will allow the researcher to determine an exposure time for optimal signal.

Stripping and Reprobing a Western blot:

1. Remove membrane from plastic following initial detection with KPL LumiGLO.
2. Rinse membrane for 30 - 60 minutes at 70°C in 2% SDS (w/v)/62.5 mM Tris-HCl (pH 6.8 at 20°C)/100 mM -mercaptoethanol.
3. Wash membrane 2 times in 10 mM Tris-HCl (pH 7.4 at 20°C)/150 mM NaCl.
4. Block membrane for 2.5 hours with KPL Detector Block or 10 mM Tris-HCl (pH 7.4 at 20°C)/150 mM NaCl/5% nonfat dry milk.
5. Repeat detection procedure.

SOUTHERN BLOTTING

Detection:

1. Perform gel electrophoresis and transfer following standard procedures.
2. Prehybridize membrane for 30 minutes to 1 hour at the appropriate hybridization temperature.
3. Add biotinylated probe to hybridization solution and hybridize 3 - 16 hours at the appropriate hybridization temperature.
4. Following hybridization perform stringency washes with SSC or SSPE following standard protocols.
5. Block membrane for 30 minutes to 1 hour with KPL Detector Block (See RELATED PRODUCTS) or other appropriate blocking solution.
6. Incubate with KPL HRP-Streptavidin (See RELATED

**KPL LumiGLO®
Chemiluminescent Substrate**

<u>Catalog No.</u>	<u>Size</u>
5430-0042 (54-61-02)	60 mL
5430-0040 (54-61-00)	240 mL
5430-0041 (54-61-01)	720 mL

TROUBLESHOOTING: BLOTTING

Problem	Corrective Measure
Excess Signal or Background	<p>time. Decrease HRP conjugate concentration.</p> <p>incubation time.</p> <p>blocking times.</p> <p>gel.</p>
No signal	<p>Verify transfer by staining protein gel with Coomassie blue or DNA gel with ethidium bromide.</p> <p>Verify protein transfer by staining membrane with Ponceau-S or Amido black.</p> <p>antibody is specific for the primary antibody.</p> <p>solutions, this will inhibit peroxidase activity.</p>
Weak signal	<p>m exposure time.</p> <p>concentration.</p> <p>incubation time.</p> <p>gel.</p> <p>has high affinity for target protein. Antibody affinity may change after denaturation of sample with SDS.</p>

ELISA PROCEDURE

All steps are at room temperature unless otherwise noted.

Note: The typical light decay of KPL LumiGLO in microtiter plates has a $t^{1/2}$ value of 60 minutes.

1. Coat an opaque white microwell plate with 100 antigen, diluted in KPL Coating Solution (See RELATED PRODUCTS) or Carbonate Buffer (pH 9.6), for 2 hours at room temperature or overnight at 2-8°C. Optimal antigen dilution must be determined experimentally.
2. Block plate for 15 - _____ ell of KPL Milk Diluent/Blocking Solution (See RELATED PRODUCTS) or other appropriate blocking solution.
3. _____ antibody diluted in blocking solution. Optimal antibody dilution must be determined experimentally.
4. Wash plate 3 times with KPL Wash Solution (See RELATED PRODUCTS) or other appropriate wash solution.
5. Incubate plate for 30 minutes to 1 hour with 100 HRP-labeled secondary antibody diluted in blocking solution or other appropriate diluent. Optimal antibody dilution must be determined experimentally.
6. Wash plate 3 times.
7. Prepare KPL LumiGLO Chemiluminescent Substrate by mixing equal volumes of Substrate A and Substrate B. Add _____ KPL LumiGLO working solution.
8. Read on a Luminometer with 1 second integration time per well. KPL LumiGLO provides consistent results when read 5 - 45 minutes after addition of substrate.

**KPL LumiGLO®
Chemiluminescent Substrate**

Catalog No.
5430-0042 (54-61-02)
5430-0040 (54-61-00)
5430-0041 (54-61-01)

Size
60 mL
240 mL
720 mL

TROUBLESHOOTING: ELISA

Problem	Corrective Measure
Excess Signal or Background	conjugate concentration. incubation times. blocking times. protein coated to plate. contaminate adjacent wells. Remove substrate from suspect well and place in another well to get a more accurate reading.
No signal	antibody is specific for the primary antibody. solutions, this will inhibit peroxidase activity. Luminometer is working correctly
Weak signal	concentration. incubation time. protein coated to the plate. has high affinity for target protein.

- Knecht, D. and Dimond, R. (1984) Visualization of Antigenic Proteins on Western Blots. *Anal. Biochem.* 136: 180-184.
- Burnette, W. (1981) Western Blotting: Electrophoretic Transfer of Proteins and Nucleic Acids From Slab Gels to Unmodified Nitrocellulose and Radiographic Detection With Antibody and Radioiodinated Protein A. *Anal. Biochem.* 102: 459-471.
- Kaufmann, et. al. (1987). The Erasable Western Blot. *Anal. Biochem.* 161: 89-95.
- Isacsson V, Wettermark, G. (1974) Chemiluminescence in Analytical Chemistry. *Anal. Chim. Acta.* 68: 339-362.
- Ausubel, R., et. al. (eds.) *Current Protocols in Molecular Biology.* John Wiley and Sons, NY.

RELATED PRODUCTS

CAT NO.

KPL Detector Block	5920-0004 (71-83-00)
KPL Wash Solution Concentrate	5150-0008 (50-63-00)
KPL HRP Streptavidin, MB grade	5950-0004 (474-3000)
KPL Biotin Wash Solution Concentrate	5960-0015 (50-63-06)
KPL 20X SSC	5960-0021 (50-86-05)
KPL Coating Solution Concentrate	5150-0014 (50-84-00)
KPL Milk Diluent/Blocking Solution	5140-0011 (50-82-01)
KPL Biodyne B Nylon Membrane	5960-0026 (60-00-51)
KPL Protein Detector LumiGLO Western Blotting Kit	5410-0009 (54-12-50)
KPL HRP Chemiluminescent Blotting Kit	5910-0027 (54-30-00)

PRODUCT SAFETY AND HANDLING

See SDS (Safety Data Sheet) for this product.

REFERENCES

- Towbin, H. et. al. (1979) Electrophoretic Transfer of Proteins From Polyacrylamide Gels to Nitrocellulose Sheets. *Proc. Natl. Acad. Sci.* 76: 4350-4354.
- Reinhart, M. and Malamud, D. (1982) Protein Transfer From Isoelectric Focusing Gels: The Native Blot. *Anal. Biochem.* 123: 229-235.
- Kricka, L. (1991) Chemiluminescent and Bioluminescent Techniques. *Clin. Chem.* 37(9): 1472-1481.

KPL LumiGLO® is a registered trademark of SeraCare Life Sciences and is protected by the following patents:

US 4598044
 Australia 575552
 Canada 1217121
 New Zealand 207095
 South Africa 84/0909
 Finland 76380
 Japan 1649482
 Belgium, Sweden, Germany, France, Netherlands, UK,
 Switzerland, Italy EPO116454